



CERTIFICATE OF CALIBRATION

MODEL NUMBER _____

SERIAL NUMBER _____

This is to certify that the instrument described above was calibrated in our facilities according to our standard procedure.

We used an ozone source with the ozone level as read on the instrument under test read simultaneously on a reference standard UV ozone analyzer whose calibration is certified to be NIST traceable. The calibration of the instrument under test is adjusted as necessary. The calibration is checked several times over several hours of testing. The calibration data is entered with the serial number, customer and date in our permanent calibration database.

This calibration can be relied upon only if our instructions included with the instrument are carefully followed. Please contact us if you have any doubts or questions.

ECO SENSORS, INC.

CALIBRATED BY _____

A handwritten signature in black ink, appearing to be "R. B.", is written over a horizontal line.

DATE OF CALIBRATION _____