



STRIPPING CHAMBER

NOTE

The stripping chamber prevents gas flow from the stripping chamber to the drain. The exit tubing must have 5 inches vertical flow to insure proper functioning of the unit. If this condition is not maintained there will be incorrect dissolved ozone readings.

Inside Diameter .375" (9.5 mm)

TO DRAIN

Approximately .15 L/min (3 gal/hr)
water flow to drain

CAUTIONS

1. Maintain 5 inches of vertical fall of the exit tubing.
2. Do not pump exiting water faster than the natural flow.

Itemref	Quantity	Title/Name, designation, material, dimension etc			Article No./Reference	
Designed by Larry Kilham	Checked by Ed Tomingas	Approved by - date Larry Kilham	Filename DOM-1	Date 5/21/05	Scale n/a	
Eco Sensors, Inc.			DOM-1 EXIT TUBING			
				900550	Edition Initial Rel	Sheet 1/1