


OZONE GENERATOR PERFORMANCE TEST

TG-40 Ozone Generator


#1

# #	Oxygen Flow in LPM	Ozone Concentration, % by weight	Ozone Production in g/hr	PSI	COMMENTS
1	5.00	6.00	25.74	20	
2	10.00	5.00	42.90	20	
3	15.00	4.25	54.70	20	
4	17.60	4.00	60.40	20	
5	20.00	3.55	60.92	20	

Ozone Concentration



Ozone Production



Test was performed at 72°F, Generator was ON for more than 1 hour.
Ozone Analyzer API 460H